

## Implant compatibility chart



Manufacturer	Implant	Titanium	GoldHue	Zirconia, white	Zirconia, shaded	Screwdriver
<b>Astra Tech</b>	<b>INTERNAL</b>					
	OsseoSpeed™/OsseoSpeed™ TX 3.0 – Yellow	x	x			0.050" Hex screwdriver
	OsseoSpeed™/OsseoSpeed™ TX 3.5, 4.0 – Aqua	x	x	x	x	
	OsseoSpeed™/OsseoSpeed™ TX 4.5, 5.0 – Lilac	x	x	x	x	
OsseoSpeed™ 4.0 – Blue (Previous)	x	x				
<b>BioHorizons</b>	<b>INTERNAL</b>					
	Internal 3.5, 4.5, 5.7	x	x	x	x	1.25 mm Hex screwdriver
	Tapered Internal 3.5, 4.5, 5.7	x	x	x	x	
	<b>EXTERNAL</b>					
	External Implant 3.5	x	x			0.050" Hex screwdriver
	External Implant 4.0, 5.0	x	x	x	x	
<b>Biomet 3i</b>	<b>INTERNAL</b>					
	Certain™ 3.25, 4.0, 4/3, 5.0, 5/4, 6.0	x	x	x	x	0.048" Hex screwdriver
	Certain™ Prevail™ 3/4/3, 4/5/4, 5/6/5, 4/3, 5/4, 6/5	x	x	x	x	
	Certain™ XP 4/5, 5/6	x	x	x	x	
	<b>EXTERNAL</b>					
	MicroMiniplant™ 3.25	x	x			0.048" Hex screwdriver
	Miniplant™ 3.25	x	x	x	x	
	Osseotite™ 3.75, 4.0, 5.0, 6.0	x	x	x	x	
Osseotite™ XP 3/4, 4/5, 5/6	x	x	x	x		
<b>Dentsply Friadent</b>	<b>INTERNAL</b>					
	XiVE S 3.0	x	x			0.048" Hex screwdriver
	XiVE S/Frialit 3.4, 3.8, 4.5, 5.5, 6.5	x	x	x	x	
<b>Keystone Dental</b>	<b>EXTERNAL</b>					
	Restore™ Small Diameter 3.3, 3.4	x	x			0.048" Hex screwdriver
	Restore™ Regular Diameter 3.75, 4.0	x	x	x	x	
	Restore™ Wide Diameter 5.0, 5.5, 6.0	x	x			
<b>Nobel Biocare</b>	<b>INTERNAL</b>					
	NobelActive™ NP 3.5	x	x	x	x	Nobel Biocare Unigrip
	NobelActive™ RP 4.3, 5.0	x	x	x	x	
	NobelReplace™ NP 3.5	x	x	x	x	
	NobelReplace™ RP 4.0	x	x	x	x	
	NobelReplace™ WP 5.0	x	x	x	x	
	NobelReplace™ 6.0	x	x	x	x	
	<b>EXTERNAL</b>					
	Brånemark System™ NP 3.3	x	x			Nobel Biocare Unigrip
	Brånemark System™ RP 3.75, 4.0	x	x	x	x	
	Brånemark System™ WP 5.0, 5.5	x	x	x	x	
	SteriOss Replace™ 3.1, 3.5, 4.3, 5.0, 6.0	x	x			0.050" Hex screwdriver
	SteriOss HL™ 3.25, 3.8, 4.5, 5.0, 6.0	x	x			
<b>Straumann</b>	<b>INTERNAL</b>					
	Bone Level 3.3 NC (Narrow CrossFit™)	x	x			Straumann screwdriver compatible
	Bone Level 4.1, 4.8 RC (Regular CrossFit™)	x	x	x	x	
	Standard/Standard Plus 4.8 RN (Regular Neck)	x	x			
	Standard/Standard Plus 4.8 WN (Wide Neck)	x	x			
	<b>EXTERNAL</b>					
Standard Plus 3.3 NN (Narrow Neck)	x	x			Straumann screwdriver compatible	
<b>Sybron Implant Solutions</b>	<b>EXTERNAL</b>					
	Endopore 3.5, 4.1, 5.0 – external	x	x			0.050" Hex screwdriver
	Entegra 3.5, 4.1, 5.0 – external	x	x			
<b>Zimmer Dental</b>	<b>INTERNAL</b>					
	Tapered Screw-Vent™ 3.5, 4.5, 5.7	x	x	x	x	1.25 mm Hex screwdriver
	Screw Vent™ 3.3, 3.7, 4.5	x	x	x	x	
	SwissPlus™ 4.8	x	x			

Available July 2010

Available July 2010

## Turnaround times

Atlantis™ Abutments is available for all major implant systems, and in titanium, gold-shaded titanium and zirconia material options. Turnaround times vary depending on material and case size.

When the case is received at the Atlantis™ production unit and the production process is started, the Atlantis™ WebOrder account will be updated with the case delivery date.

Standard turnaround times are as follows:

Atlantis™ Abutment fabrication time (days)			
Case size	Titanium	Gold-shaded titanium	Zirconia white and shaded
1–4 unit cases	2	5	5
5+ unit cases	3	6	6

### Determine return date

1. Select your choice of delivery method and calculate the number of days it will take to ship the case from the laboratory to the Atlantis™ production unit. Astra Tech DHL account, # 960441255 can be used and the cost will be charged on the invoice.
2. Add on the days the case will be in-house at the Atlantis™ production unit. Calculation of fabrication time starts when the case is received at the Atlantis™ production unit before noon.
3. Astra Tech will ship the case free of charge via next day delivery.

### Note:

- Design approvals to process case must be granted by midnight.
- Material, on-hold and case information issues must be resolved by midnight.
- Weekends and holidays are excluded.

### Example

*1–4 unit cases in titanium, received at Atlantis™ production unit before noon Monday will be shipped Wednesday and received in the laboratory Thursday.*

### Additional information

Atlantis™ Abutments is delivered with the corresponding abutment screw at no extra charge. The implant manufacturer's recommended torque should always be used when placing the abutment screw.

For the most current version of the implant compatibility chart and turnaround times go to [www.atlantisabutment.com](http://www.atlantisabutment.com)

